



CURRENT TECHNOLOGY IN HIP REPLACEMENT

Sholahuddin Rhatomy¹, Asep Santoso², Krisna Yuarno Phatama³, Nicolaas C. Budhiparama⁴

¹Department of Orthopedics and Traumatology, Dr. Soeradji Tirtonegoro General Hospital, Klaten, Indonesia

^{1,4}Faculty of Medicine, Public Health, and Nursing, Universitas Gadjah Mada, Yogyakarta, Indonesia

²Department of Orthopaedic and Traumatology, Prof. Dr. R. Soeharso Orthopaedic Hospital, Universitas Sebelas Maret, Surakarta, Indonesia

³Department of Orthopaedic and Traumatology, Dr. Saiful Anwar General Hospital, Universitas Brawijaya, Malang, Indonesia

⁴Nicolaas Institute of Constructive Orthopaedics. Research and Education Foundation for Arthroplasty & Sports Medicine, Medistra Hospital, Jakarta, Indonesia

⁴Department of Orthopaedic and Traumatology, Universitas Airlangga, Surabaya, Indonesia

⁴Department of Orthopaedic Surgery, Leiden University Medical Center, Leiden, The Netherlands

⁴Faculty of Vocational Studies, Universitas Airlangga, Surabaya, Indonesia

INTRODUCTION

Hip replacement, also known as total hip arthroplasty (THA), is widely regarded as an incredibly effective surgical procedure, often hailed as a ground-breaking achievement in medicine (1). The initial development of total hip replacement dates to 1938 when Wiles made significant progress (2). However, it wasn't until the 1960s that this procedure gained widespread recognition and popularity. Sir John Charnley's introduction of "low-friction arthroplasty" revolutionized the treatment of arthritic joints (1). Over time, significant progress has been made in THA's design, materials, and surgical techniques, significantly enhancing patient satisfaction, minimizing surgical complications, and improving clinical outcomes. Patients' expectations regarding life after total hip arthroplasty (THA) have significantly shifted. They prioritize long-term survival and the various aspects of maintaining a good quality of life. They have a strong desire to pursue their professional

and personal interests, which require a high level of physical activity (3).

The initial technological advancement in hip replacement surgery was the introduction of Minimally Invasive Surgery (MIS), which reduced the surgical footprint. MIS techniques involve smaller incisions and less soft tissue disruption, reducing postoperative pain and faster recovery times. This approach contrasts with traditional open surgery, which typically requires larger incisions and more extensive muscle dissection. MIS procedures have been proven to enhance cosmetic outcomes, minimize hospital stays, and accelerate rehabilitation (4,5).

A growing focus has recently been on customizing joint replacement procedures to suit individual patients rather than relying on a generic and uniform approach. A computer-assisted surgery (CAS) and robotic-assisted surgery represent significant technological leaps in hip replacement. CAS utilizes preoperative imaging and intraoperative navigation to

enhance the precision of implant placement. CAS improves the alignment of the acetabular and femoral components, thereby reducing the risk of dislocation and improving functional outcomes (6). Robotic-assisted systems, such as the MAKO robotic arm, have refined this precision. These systems allow for patient-specific surgical planning and real-time adjustments during the procedure. Studies have shown that robotic-assisted surgery results in superior implant positioning and reduced variability in component placement (6,7). The integration of these technologies into clinical practice represents a significant advancement in achieving optimal surgical outcomes.

Another enhancement in hip replacement surgery is the development of advanced biomaterials, which have been pivotal in extending the longevity of hip implants. Traditional metal-on-polyethylene bearings, while effective, are prone to wear and osteolysis over time. Recent advancements have introduced ceramic-on-ceramic and ceramic-on-polyethylene bearings, which exhibit lower wear rates and higher biocompatibility (8). Ceramic materials, being more complex and smoother than metals, significantly reduce friction and wear particles, enhancing the implant's lifespan. Highly porous metals, such as trabecular metal, have also been developed to improve osseointegration. These materials mimic the trabecular structure of natural bone, promoting biological fixation and reducing the reliance on bone cement. The benefits of trabecular metal are in providing a stable and durable implant-bone interface, which is crucial for

the long-term success of hip replacements (9).

Enhanced Recovery After Surgery (ERAS) protocols represent a multidisciplinary approach to optimizing perioperative care. These protocols focus on preoperative education, optimized pain management, and early mobilization. ERAS protocols have been shown to reduce hospital stays, decrease complications, and improve overall patient outcomes in hip replacement surgery. Implementing ERAS has led to a paradigm shift in postoperative care, emphasizing patient-centered approaches and evidence-based practices (10).

Despite these advancements, challenges remain in hip replacement surgery. Implant longevity in younger, more active patients continues to be a concern. Additionally, periprosthetic joint infections (PJIs) pose a significant risk to implant success. Current research explores antibacterial coatings and systemic antibiotic regimens to address PJIs (11). Integrating artificial intelligence (AI) and machine learning (ML) into hip replacement surgery offers promising future directions. AI-driven predictive analytics can assist in preoperative planning by identifying patients at higher risk of complications and customizing treatment plans. Furthermore, ML algorithms can analyze postoperative data to monitor patient recovery and predict long-term outcomes (12)

REFERENCE

1. Learmonth ID, Young C, Rorabeck C. The operation of the century: total hip replacement. *Lancet*.

- 2007;370(9597):1508-1519.
doi:10.1016/S0140-6736(07)60457-7
2. WILES P. The surgery of the osteoarthritic hip. *Br J Surg.* 1958;45(193):488-497.
doi:10.1002/bjs.18004519315.
 3. Puliero B, Blakeney WG, Beaulieu Y, Vendittoli PA. Joint Perception After Total Hip Arthroplasty and the Forgotten Joint. *J Arthroplasty.* 2019;34(1):65-70.
doi:10.1016/j.arth.2018.09.086
 4. Ramadanov N, Marinova-Kichikova P, Hable R, Dimitrov D. Minimally Invasive versus Conventional Approaches in Total Hip Arthroplasty: A Systematic Review and Meta-Analysis of 47 Randomized Controlled Trials. *Prosthesis.* 2023; 5(3):962-991.
doi:10.3390/prosthesis5030067
 5. Bal BS, Vallurupalli S. A modified two-incision technique for primary total hip arthroplasty. *Indian J Orthop.* 2008;42(3):267-274.
doi:10.4103/0019-5413.41850
 6. Xu K, Li YM, Zhang HF, Wang CG, Xu YQ, Li ZJ. Computer navigation in total hip arthroplasty: a meta-analysis of randomized controlled trials. *Int J Surg.* 2014;12(5):528-533.
doi:10.1016/j.ijisu.2014.02.014
 7. Kim K, Kwon S, Kwon J, Hwang J. A review of robotic-assisted total hip arthroplasty. *Biomed Eng Lett.* 2023;13(4):523-535.
doi:10.1007/s13534-023-00312-9
 8. Malchau H, Herberts P, Eisler T, Garellick G, Söderman P. The Swedish Total Hip Replacement Register [published correction appears in *J Bone Joint Surg Am.* 2004 Feb;86-A(2):363]. *J Bone Joint Surg Am.* 2002;84-A Suppl 2:2-20.
doi:10.2106/00004623-200200002-00002
 9. Tanzer M, Chuang PJ, Ngo CG, Song L, TenHuisen KS. Characterization of bone ingrowth and interface mechanics of a new porous 3D printed biomaterial. *Bone Joint J.* 2019;101-B(6_Supple_B):62-67.
doi:10.1302/0301-620X.101B6.BJJ-2018-1472.R1
 10. Ljungqvist O, Scott M, Fearon KC. Enhanced Recovery After Surgery: A Review. *JAMA Surg.* 2017;152(3):292-298.
doi:10.1001/jamasurg.2016.4952
 11. Xie H, Liu Y, An H, et al. Recent advances in prevention, detection and treatment in prosthetic joint infections of bioactive materials. *Front Bioeng Biotechnol.* 2022;10:1053399.
doi:10.3389/fbioe.2022.1053399
 12. Lisacek-Kiosoglous AB, Powling AS, Fontalis A, Gabr A, Mazomenos E, Haddad FS. Artificial intelligence in orthopaedic surgery. *Bone Joint Res.* 2023;12(7):447-454.
doi:10.1302/2046-3758.127.BJR-2023-0111.R1