

Analysis of Wear in UHMWPE Artificial Hip Joint Using Finite Element Method: A Review

Mujib Wahyudi ¹, Achmad Zain Nur ²

¹ Departement of Mechanical Engineering, University of Trunojoyo Madura, Jl. Raya Telang PO. Box 2 Kamal Bangkalan 69161, Indonesia

² Departement of Electrical Engineering, University of Trunojoyo Madura, Jl. Raya Telang PO. Box 2 Kamal Bangkalan 69161, Indonesia

E-mail: ¹ mujib.wahyudi@trunojoyo.ac.id

DOI : <https://doi.org/10.52620/sainsdata.v2i2.131>

Abstract

Ultra-high-molecular-weight polyethylene (UHMWPE) is one of the most commonly used materials in the fabrication of artificial hip joints due to its excellent mechanical properties, biocompatibility, and wear resistance. Despite its advantages, the wear of UHMWPE components over time leads to the generation of wear particles, which are responsible for osteolysis and implant failure. The Finite Element Method (FEM) has emerged as a powerful tool to simulate the wear mechanisms of UHMWPE in hip implants, providing a deeper understanding of the stress distribution, wear prediction, and optimization of implant designs. This review explores the current state of FEM applications in UHMWPE wear analysis, focusing on the material properties, wear mechanisms, FEM simulation models, and future directions in this research area.

Keywords: Artificial Hip Joint, Finite Element Method, Implant Design Optimization, UHMWPE, Wear Analysis, Wear Mechanisms

1. Introduction

Wear of UHMWPE in artificial hip joints remains one of the most significant factors limiting the longevity of implants, leading to revision surgeries due to osteolysis and failure of the prosthesis [1], [2], [3], [4]. The use of FEM in the analysis of UHMWPE wear has increased over recent years, with several studies demonstrating its ability to predict wear behavior under a variety of loading and kinematic conditions [5], [6]. FEM simulations offer insights into the intricate interactions between implant components and surrounding tissues, enabling the optimization of material choices, implant designs, and surgical procedures [7].

The understanding of wear mechanisms such as adhesive, abrasive, and fatigue wear in UHMWPE components is vital in designing more durable implants [8], [9]. Moreover, FEM models provide a virtual environment where various parameters, such as loading conditions, implant geometry, and material behavior, can be systematically varied and analyzed [10], [11]. Figure 1 presents a comparison of predicted volumetric wear rates of ultra-high molecular weight polyethylene (UHMWPE) under different loading conditions, as derived from the computational model and validated against hip simulator test results. The graph illustrates that the wear rates predicted by the modified frictional work-dependent equation closely align with the experimental data, demonstrating improved accuracy over previous models. Notably, the wear rates increase with varying contact pressures, highlighting the significant impact of loading conditions on material degradation. This figure underscores the effectiveness of incorporating both frictional work and contact pressure in enhancing the predictive capabilities of wear modeling for total hip joint replacements.



This work is licensed under a [Creative Commons Attribution 4.0 International License](https://creativecommons.org/licenses/by/4.0/)

©Author (s)




Figure 1. Comparison of predicted volumetric wear rates averaged over 5 million gait cycles between the computational models using frictional work dependent and simulator test (Liu et al., 2021).

This review aims to present a thorough examination of FEM-based approaches used to study UHMWPE wear, providing a comprehensive overview of the material properties, simulation models, and the future trajectory of wear analysis for artificial hip joints.

2. Material Properties of UHMWPE

Ultra-high-molecular-weight polyethylene (UHMWPE) is widely utilized in joint prostheses due to its combination of favorable properties such as high molecular weight, superior toughness, low friction, and excellent wear resistance. The intrinsic properties of UHMWPE allow it to withstand high mechanical loads, which are particularly important for joint applications. However, the wear of UHMWPE, particularly under cyclic loading conditions, can lead to the generation of wear debris. These particles trigger an inflammatory response in the surrounding tissue, leading to the activation of osteoclasts (bone-resorbing cells) and subsequent bone loss around the implant. We know this process as osteolysis. [12].

Fig. 2 depict the radiographic changes in osteolytic zones following head and inlay revision surgery. They illustrate a significant reduction in the size of these zones, with a noted decrease of 46% within a follow-up period of 33 months. The figures highlight the effectiveness of the revision surgery in managing osteolysis, showcasing the potential for improved outcomes in patients with large radiolytic areas. Overall, these visual representations support the study's findings that revision surgery can halt or even reverse the progression of osteolytic processes around the femoral component.



Figure 2. Reduction of osteolytic zones by 46% within 33 months [12].

Recent advancements in UHMWPE material science focus on enhancing its wear resistance through various strategies, such as cross-linking and the incorporation of additives. Cross-linking, either through gamma radiation or other techniques, has been found to significantly reduce wear rates by increasing the material's resistance to plastic deformation [13], [14], [15], [16]. However, the increased brittleness and reduced fatigue resistance of cross-linked UHMWPE have led to concerns regarding its long-term mechanical integrity [14].

Additionally, the integration of antioxidants, such as vitamin E, into UHMWPE has been shown to improve oxidative stability and reduce wear rates under oxidative degradation conditions [15]. Figure 3 presents a comparative analysis of the wear rates of three types of Ultra-High Molecular Weight Polyethylene (UHMWPE) samples: conventional UHMWPE, highly cross-linked UHMWPE (HXL-UHMWPE), and Vitamin E-blended HXL-UHMWPE. The graph illustrates the mass loss (in mm^3) per 0.1 million cycles during wear testing conducted using a hip joint simulator.

The results indicate that conventional UHMWPE experiences the highest mass loss, reflecting its greater susceptibility to wear compared to the other two materials. In contrast, HXL-UHMWPE demonstrates a significant reduction in mass loss, highlighting the enhanced wear resistance attributed to the cross-linking process. Notably, Vitamin E-blended HXL-UHMWPE exhibits the lowest mass loss among the three, suggesting that the incorporation of Vitamin E not only improves wear resistance but also effectively reduces oxidative degradation and inhibits fatigue crack propagation. This data underscores the potential of both cross-linking and Vitamin E blending in enhancing the biotribological performance of UHMWPE for artificial joint applications, making it a promising material for improving the longevity and functionality of hip implants.




Figure 3. Wear rate of three types of UHMWPE [16].

Incorporating carbon-based nanomaterials into UHMWPE has also been studied to further enhance its mechanical and tribological properties [17]. These innovations are aimed at mitigating wear-induced complications in hip replacements, extending implant lifetimes, and improving patient outcomes.

3. Wear Mechanisms in UHMWPE

The wear mechanisms in UHMWPE are primarily driven by the joint's complex loading conditions, which result in different types of wear: adhesive, abrasive, and fatigue wear. Adhesive wear occurs when there is a transfer of material between the articulating surfaces due to high contact stresses, leading to the formation of wear particles [18]. Abrasive wear occurs when foreign particles or rough surfaces scratch and remove material from the UHMWPE surface [19], [20]. Fatigue wear, on the other hand, is associated with the cyclic loading of the material, which leads to the gradual formation of cracks and material loss [21], [22].

The contribution of these wear mechanisms to the overall wear rate depends on several factors, such as the type of loading (e.g., walking vs. running), the surface roughness of the components, and the lubricating conditions. FEM-based simulations have proven useful in quantifying the distribution of stresses across the surface of the UHMWPE, aiding in the identification of regions with the highest wear potential. By simulating various loading conditions, FEM can predict the wear behavior under normal and extreme conditions, providing valuable insights into how wear initiates and propagates over time [10], [23].

4. Finite Element Analysis of UHMWPE Wear

Finite Element Analysis (FEA) of UHMWPE wear provides a computational framework to simulate and predict material degradation in artificial hip joints under varying mechanical loads and motion cycles. By incorporating wear laws and material properties specific to UHMWPE, FEA enables detailed insights into the effects of stress distribution, contact pressures, and kinematics on the wear behavior of the polymer. This analytical approach aids in optimizing implant design, improving material selection, and reducing the risk of failure due to wear, ultimately enhancing the longevity of artificial hip joints.

Avoid hyphenation at the end of a line. Symbols denoting vectors and matrices should be indicated in bold type. Scalar variable names should normally be expressed using italics. Weights and measures should be expressed in SI units. All non-standard abbreviations or symbols must be defined when first mentioned, or a glossary provided.

4.1. Contact Mechanics

The contact pressure and contact area between the femoral head and acetabular cup in hip joint prostheses are critical factors in determining wear behavior. FEM models allow for the calculation of contact pressures and the distribution of stresses under different loading conditions [24], [25]. These simulations have demonstrated that the contact area, joint alignment, and head size significantly influence the wear rate of UHMWPE. For example, larger femoral heads tend to reduce peak contact stresses but may increase volumetric wear due to increased sliding distances [26], [27], [28].

Figure 4 presents the evolution of wear patterns in the XLPE (cross-linked polyethylene) bearing liner over a simulated duration of 10 million cycles, which approximates to 10 years of walking activity. The figure compares the maximum wear depth of the XLPE bearing liner for femoral head sizes of 22 mm and 36 mm. Notably, the results indicate that as the femoral head diameter increases, the maximum linear wear depth decreases. Specifically, the 22 mm femoral head demonstrates a maximum linear wear of 0.37 mm, whereas the 36 mm femoral head exhibits a reduced maximum linear wear of 0.22 mm. These findings suggest that larger femoral heads may contribute to decreased wear on the bearing liner, highlighting a potential advantage in utilizing larger head sizes to minimize wear in total hip arthroplasty. The overall wear pattern remains consistent throughout the analysis, reinforcing the relationship between femoral head size and wear evolution in hip prostheses.




Figure 4. Evolution of wear pattern over 10 years at XLPE bearing liner for different head sizes [28].

Patient-specific anatomical data, such as bone morphology and gait analysis, are increasingly incorporated into FEM models to create more accurate predictions of wear behavior. These models help optimize implant geometry and alignment to minimize wear and extend the service life of the prosthesis [29], [30], [31], [32].

4.2. Wear Simulation Models

Wear prediction models typically rely on Archard's wear law, which correlates the wear volume to the normal load, sliding distance, and wear coefficient [23], [33], [34]. Equation (1) represents Archard's wear law, which is utilized to model adhesive wear in tribological systems [35]. The equation is expressed as follows:

$$Volume\ Rate = K_{wear} = \frac{F V_{rel}}{H_{cup}} \quad (1)$$

K_{wear} denotes Archard's wear constant, a material-specific parameter that quantifies the wear rate. F represents the normal load applied to the contact surfaces. V_{rel} indicates the relative sliding velocity between the two surfaces in contact. H_{cup} refers to the hardness of the acetabular cup material.

FEM models integrate this law with material-specific properties to simulate wear progression over time. The addition of time-dependent properties, such as the change in material behavior with wear, further enhances the accuracy of these models [35].

Several studies have focused on developing hybrid models that combine FEM simulations with experimental validation, such as wear testing and mechanical testing. These models incorporate various loading conditions, lubrication effects, and even patient-specific data, providing more accurate predictions of wear behavior under real-world conditions [10].

4.3. Multidirectional Wear

Wear in hip joint implants is often multidirectional due to the complex motion of the femoral head and acetabular cup during normal joint activity. FEM-based simulations have been used to capture these multidirectional wear patterns, which are difficult to replicate in simple laboratory tests [36]. These simulations provide insights into how the direction and magnitude of loading affect the wear progression in the UHMWPE component, allowing for the development of more accurate wear prediction models [23].

4.4. Fatigue and Frictional Effects

Fatigue wear is a major concern for the long-term performance of UHMWPE, as repeated loading cycles can cause material failure. FEM has been instrumental in predicting fatigue failure by simulating the material's response to cyclic loading conditions [37]. Additionally, frictional heat generated during articulation plays a role in the wear process, with studies showing that higher friction leads to increased wear rates [38].

5. Applications of FEM in Design Optimization

FEM has become an essential tool in optimizing the design of hip joint prostheses to reduce wear and improve patient outcomes. By simulating different implant geometries, head sizes, and material choices, FEM models can help identify designs that minimize wear while maintaining mechanical stability [34], [39], [40], [41], [42]. Additionally, FEM can be used to evaluate the effects of surgical factors, such as implant orientation and alignment, on wear behavior [43], [44], [45]. Figure 5 presents the new hip joint prosthesis design features uni-directional articulations that allow for controlled sliding motion, contrasting with the multi-directional movement of conventional ball-and-socket implants. It incorporates an interlocking mechanism at the cup-flexor and flexor-rotator articulations, simplifying assembly and enhancing stability during movement. The design eliminates protrusions and grooves on the articulating surfaces, which reduces wear debris and improves tribological performance. Additionally, the new design achieves a greater range of motion in flexion/extension compared to dual-mobility implants, while still meeting ISO standards for hip joint functionality.




Figure 5. The hip implant developed in this work, showing (a) assembly view, and; (b) an exploded view with the components adopted nomenclature [34].

Emerging techniques, such as additive manufacturing, have also benefited from FEM simulations, as they allow for the production of patient-specific implants with customized geometries [46], [47], [48], [49]. This approach can further reduce wear by optimizing the fit and function of the implant to the patient's unique anatomy.

6. Challenges and Future Directions

Despite the progress made in FEM simulations of UHMWPE wear, several challenges remain. One of the most significant challenges is the accurate incorporation of patient-specific data, such as bone structure and gait, into wear prediction models. Although FEM provides valuable insights, the real-world complexity of wear is difficult to capture fully in a computational model. Moreover, the

integration of third-body wear, where foreign particles accelerate wear, and tribochemical effects remains an area requiring further research [50].

In the future, advancements in machine learning and artificial intelligence may offer new ways to enhance FEM simulations. These technologies can potentially process large datasets and optimize models based on real-world clinical data, leading to more accurate predictions of implant performance and wear progression [48], [51], [52]. Additionally, improving the modeling of nanoscale interactions between the wear debris and UHMWPE will provide a more detailed understanding of wear mechanisms at the microscopic level [53].

7. Conclusions

The use of FEM in analyzing the wear of UHMWPE artificial hip joints has proven to be a powerful tool for improving the design, material selection, and optimization of hip implants. By accurately simulating wear mechanisms under a variety of conditions, FEM has provided valuable insights into how wear initiates and propagates over time. While challenges remain, particularly in integrating patient-specific data and third-body wear, future advancements in computational modelling and material science are expected to further enhance the longevity and performance of UHMWPE implants.

References

- [1] S. B. Goodman and J. Gallo, "Periprosthetic osteolysis: Mechanisms, prevention and treatment," Dec. 01, 2019, *MDPI*. doi: 10.3390/jcm8122091.
- [2] A. M. Kandahari, X. Yang, K. A. Laroche, A. S. Dighe, D. Pan, and Q. Cui, "A review of UHMWPE wear-induced osteolysis: The role for early detection of the immune response," Jul. 12, 2016, *Sichuan University*. doi: 10.1038/boneres.2016.14.
- [3] L. Mattei, F. Di Puccio, B. Piccigallo, and E. Ciulli, "Lubrication and wear modelling of artificial hip joints: A review," *Tribol Int*, vol. 44, no. 5, pp. 532–549, May 2011, doi: 10.1016/j.triboint.2010.06.010.
- [4] L. Mattei, F. Di Puccio, and E. Ciulli, "A comparative study of wear laws for soft-on-hard hip implants using a mathematical wear model," *Tribol Int*, vol. 63, pp. 66–77, 2013, doi: 10.1016/j.triboint.2012.03.002.
- [5] Y. Kumaran, J. M. Bonsu, S. Tripathi, S. M. Soehnlén, and C. E. Quatman, "Phase-specific changes in hip joint loading during gait following sacroiliac joint fusion: Findings from a finite element analysis," *Clinical Biomechanics*, vol. 122, Feb. 2025, doi: 10.1016/j.clinbiomech.2025.106429.
- [6] F. Jahani *et al.*, "Importance of dynamics in the finite element prediction of plastic damage of polyethylene acetabular liners under edge loading conditions," *Med Eng Phys*, vol. 95, pp. 97–103, Sep. 2021, doi: 10.1016/j.medengphy.2021.07.010.
- [7] C. Luo *et al.*, "Femoral Stress Changes after Total Hip Arthroplasty with the Ribbed Prosthesis: A Finite Element Analysis," *Biomed Res Int*, vol. 2020, 2020, doi: 10.1155/2020/6783936.

- [8] V. Saikko, O. Morad, and R. Viitala, "Friction RandomPOD—A new method for friction measurement in noncyclic, multidirectional, dynamic pin-on-disk tests for orthopaedic bearing materials," *J Biomech*, vol. 118, Mar. 2021, doi: 10.1016/j.jbiomech.2021.110273.
- [9] R. I. Taylor, "Rough Surface Contact Modelling—A Review," May 01, 2022, *MDPI*. doi: 10.3390/lubricants10050098.
- [10] F. Liu, Y. He, Z. Gao, and D. Jiao, "Enhanced computational modelling of UHMWPE wear in total hip joint replacements: The role of frictional work and contact pressure," *Wear*, vol. 482–483, Oct. 2021, doi: 10.1016/j.wear.2021.203985.
- [11] L. Wang, G. Isaac, R. Wilcox, A. Jones, and J. Thompson, "Finite element analysis of polyethylene wear in total hip replacement: A literature review," Nov. 01, 2019, *SAGE Publications Ltd*. doi: 10.1177/0954411919872630.
- [12] L. Pisecky, G. Hipmair, B. Schauer, and N. Böhler, "Osteolysis in total hip arthroplasty after head and inlay revision surgery," *J Orthop*, vol. 14, no. 1, pp. 192–194, Mar. 2017, doi: 10.1016/j.jor.2016.12.004.
- [13] S. M. Kurtz and J. D. Patel, *6 The Clinical Performance of Highly Cross-linked UHMWPE in Hip Replacements*, Third Edit. Elsevier Inc., 2016. doi: 10.1016/B978-0-323-35401-1/00006-5.
- [14] L. Xiong and D. Xiong, "The influence of irradiation dose on mechanical properties and wear resistance of molded and extruded ultra high molecular weight polyethylene," *J Mech Behav Biomed Mater*, vol. 9, pp. 73–82, 2012, doi: 10.1016/j.jmbbm.2012.01.002.
- [15] J. Schwiesau, B. Fritz, G. Bergmann, A. L. Puente Reyna, C. Schilling, and T. M. Grupp, "Influence of radiation conditions on the wear behaviour of Vitamin E treated UHMWPE gliding components for total knee arthroplasty after extended artificial aging and simulated daily patient activities," *J Mech Behav Biomed Mater*, vol. 122, Oct. 2021, doi: 10.1016/j.jmbbm.2021.104652.
- [16] G. Chen, Z. Ni, S. Qian, and Y. Zhao, "Biotribological behaviour of Vitamin E-blended highly cross-linked UHMWPE in a hip joint simulator," *Industrial Lubrication and Tribology*, vol. 68, no. 5, pp. 548–553, 2016, doi: 10.1108/ILT-09-2015-0130.
- [17] S. Yousef, A. Visco, G. Galtieri, D. Nocita, and C. Espro, "Wear behaviour of UHMWPE reinforced by carbon nano fi ller and paraf fi n oil for joint replacement," *Material Science and Engineering C*, vol. 73, pp. 234–244, 2017, doi: 10.1016/j.msec.2016.11.088.
- [18] M. C. Galetz, S. H. Seiferth, B. Theile, and U. Glatzel, "Potential for adhesive wear in friction couples of UHMWPE running against oxidized zirconium, titanium nitride coatings, and cobalt-chromium alloys," *J Biomed Mater Res B Appl Biomater*, vol. 93, no. 2, pp. 468–475, May 2010, doi: 10.1002/jbm.b.31604.
- [19] S. Sharma, J. Bijwe, S. Panier, and M. Sharma, "Abrasive wear performance of SiC-UHMWPE nano-composites - Influence of amount and size," *Wear*, vol. 332–333, pp. 863–871, May 2015, doi: 10.1016/j.wear.2015.01.012.

-
- [20] A. de A. Lucas, J. D. Ambrósio, H. Otaguro, L. C. Costa, and J. A. M. Agnelli, "Abrasive wear of HDPE/UHMWPE blends," *Wear*, vol. 270, no. 9–10, pp. 576–583, Apr. 2011, doi: 10.1016/j.wear.2011.01.011.
- [21] F. Ansari, M. D. Ries, and L. Pruitt, "Effect of processing, sterilization and crosslinking on UHMWPE fatigue fracture and fatigue wear mechanisms in joint arthroplasty," Jan. 01, 2016, *Elsevier Ltd*. doi: 10.1016/j.jmbbm.2015.08.026.
- [22] E. Oral, K. K. Wannomae, N. Hawkins, W. H. Harris, and O. K. Muratoglu, " α -Tocopherol-doped irradiated UHMWPE for high fatigue resistance and low wear," *Biomaterials*, vol. 25, no. 24, pp. 5515–5522, Nov. 2004, doi: 10.1016/j.biomaterials.2003.12.048.
- [23] V. Jangid, A. K. Singh, and A. Mishra, "Wear Simulation of Artificial Hip Joints: Effect of Materials," 2019. [Online]. Available: www.sciencedirect.comwww.materialstoday.com/proceedings2214-7853
- [24] S. Shankar, L. Prakash, and M. Kalayarasan, "Finite element analysis of different contact bearing couples for human hip prosthesis," 2013.
- [25] S. Shankar, R. Nithyaprakash, B. R. Santhosh, M. S. Uddin, and A. Pramanik, "Finite element submodeling technique to analyze the contact pressure and wear of hard bearing couples in hip prosthesis," *Comput Methods Biomech Biomed Engin*, vol. 23, no. 8, pp. 422–431, Jun. 2020, doi: 10.1080/10255842.2020.1734794.
- [26] M. Al-Hajjar, J. Fisher, S. Williams, J. L. Tipper, and L. M. Jennings, "Effect of femoral head size on the wear of metal on metal bearings in total hip replacements under adverse edge-loading conditions," *J Biomed Mater Res B Appl Biomater*, vol. 101 B, no. 2, pp. 213–222, 2013, doi: 10.1002/jbm.b.32824.
- [27] M. B. Cross, D. Nam, and D. J. Mayman, "Ideal Femoral Head Size in Total Hip Arthroplasty Balances Stability and Volumetric Wear," *HSS Journal*, vol. 8, no. 3, pp. 270–274, Oct. 2012, doi: 10.1007/s11420-012-9287-7.
- [28] A. Ashkanfar, S. M. S. Toh, R. English, D. J. Langton, and T. J. Joyce, "The impact of femoral head size on the wear evolution at contacting surfaces of total hip prostheses: A finite element analysis," *J Mech Behav Biomed Mater*, vol. 153, May 2024, doi: 10.1016/j.jmbbm.2024.106474.
- [29] O. L. A. Harrysson, Y. A. Hosni, and J. F. Nayfeh, "Custom-designed orthopedic implants evaluated using finite element analysis of patient-specific computed tomography data: Femoral-component case study," *BMC Musculoskelet Disord*, vol. 8, 2007, doi: 10.1186/1471-2474-8-91.
- [30] O. Abdelaal, S. Darwish, H. El-Hofy, and Y. Saito, "Patient-specific design process and evaluation of a hip prosthesis femoral stem," *International Journal of Artificial Organs*, vol. 42, no. 6, pp. 271–290, Jun. 2019, doi: 10.1177/0391398818815479.
- [31] Y. Jun and K. Choi, "Design of patient-specific hip implants based on the 3D geometry of the human femur," *Advances in Engineering Software*, vol. 41, no. 4, pp. 537–547, 2010, doi: 10.1016/j.advengsoft.2009.10.016.

- [32] D. Scherb *et al.*, "Optimizing the individual design of revision hip implants – evaluating the effects for bone and stem using patient-specific finite element models," in *DS 133: Proceedings of the 35th Symposium Design for X (DFX2024)*, The Design Society, 2024, pp. 015–024. doi: 10.35199/dfx2024.02.
- [33] A. Ruggiero and A. Sicilia, "Lubrication modeling and wear calculation in artificial hip joint during the gait," *Tribol Int*, vol. 142, Feb. 2020, doi: 10.1016/j.triboint.2019.105993.
- [34] D. Dalli, J. Buhagiar, P. Mollicone, and P. Schembri Wismayer, "A novel hip joint prosthesis with uni-directional articulations for reduced wear," *J Mech Behav Biomed Mater*, vol. 127, Mar. 2022, doi: 10.1016/j.jmbbm.2021.105072.
- [35] G. Srivastava, N. Christian, and C. Fred Higgs, "A predictive framework of the tribological impact of physical activities on metal-on-plastic hip implants," *Biotribology*, vol. 25, Mar. 2021, doi: 10.1016/j.biotri.2020.100156.
- [36] J. E. Abu Qudeiri, A. Abdudeen, M. R. Sahadevan, and A. Padmanabhan M, "Numerical investigation on the wear characteristics of hip implant under static loading," *Heliyon*, vol. 10, no. 4, Feb. 2024, doi: 10.1016/j.heliyon.2024.e26151.
- [37] J. Reginald, M. Kalayarasan, K. N. Chethan, and P. Dhanabal, "Static, dynamic, and fatigue life investigation of a hip prosthesis for walking gait using finite element analysis," *International Journal of Modelling and Simulation*, vol. 43, no. 5, pp. 797–811, 2023, doi: 10.1080/02286203.2023.2212346.
- [38] D. Nečas, H. Usami, T. Niimi, Y. Sawae, I. Křupka, and M. Hartl, "Running-in friction of hip joint replacements can be significantly reduced: The effect of surface-textured acetabular cup," *Friction*, vol. 8, no. 6, pp. 1137–1152, Dec. 2020, doi: 10.1007/s40544-019-0351-x.
- [39] J. Xu, K. Wang, M. Gao, Z. Tu, S. Zhang, and J. Tan, "Biomechanical performance design of joint prosthesis for medical rehabilitation via generative structure optimization," *Comput Methods Biomech Biomed Engin*, pp. 1163–1179, 2020, doi: 10.1080/10255842.2020.1789970.
- [40] A. Ait Moussa, J. Fischer, R. Yadav, and M. Khandaker, "Minimizing Stress Shielding and Cement Damage in Cemented Femoral Component of a Hip Prosthesis through Computational Design Optimization," *Adv Orthop*, vol. 2017, 2017, doi: 10.1155/2017/8437956.
- [41] Rufus. Fraanje and Hidde. Duivenvoorden, *Proceedings of the 2018 19th International Conference on Research and Education in Mechatronics (REM 2018) : the Hague University of Applied Sciences, Delft, the Netherlands, June 7-8, 2018*. IEEE, 2018.
- [42] L. Fanton *et al.*, "Hip joint replacement based on linear cylindrical articulations for reduced wear: A radical change in design," *Invention Disclosure*, vol. 4, p. 100030, 2024, doi: 10.1016/j.inv.2024.100030.
- [43] A. Ruggiero, M. Merola, and S. Affatato, "Finite element simulations of hard-on-soft hip joint prosthesis accounting for dynamic loads calculated from a Musculoskeletal model during walking," *Materials*, vol. 11, no. 4, Apr. 2018, doi: 10.3390/ma11040574.
- [44] C. González-Bravo, M. A. Ortega, J. Buján, B. De La Torre, and L. Barrios, "Personalized Medicine Wear Risk Prevention and Reduction in Total Hip Arthroplasty. A Personalized

- Study Comparing Cement and Cementless Fixation Techniques Employing Finite Element Analysis," *J. Pers. Med*, p. 780, 2021, doi: 10.3390/jpm.
- [45] Y. T. Lin, J. S. S. Wu, and J. H. Chen, "The study of wear behaviors on abducted hip joint prostheses by an alternate finite element approach," *Comput Methods Programs Biomed*, vol. 131, pp. 143–155, Jul. 2016, doi: 10.1016/j.cmpb.2016.04.015.
- [46] A. Namvar *et al.*, "Finite element analysis of patient-specific additive-manufactured implants," *Front Bioeng Biotechnol*, vol. 12, 2024, doi: 10.3389/fbioe.2024.1386816.
- [47] L. Maslov *et al.*, "Some aspects of custom-made 3d-printed hip joint implant structural simulation," in *IOP Conference Series: Materials Science and Engineering*, IOP Publishing Ltd, Dec. 2020. doi: 10.1088/1757-899X/986/1/012035.
- [48] D. Milone, D. D'Andrea, and D. Santonocito, "Smart Design of Hip Replacement Prostheses Using Additive Manufacturing and Machine Learning Techniques," *Prosthesis*, vol. 6, no. 1, pp. 24–40, Feb. 2024, doi: 10.3390/prosthesis6010002.
- [49] L. Risse, S. Woodcock, J. P. Brüggemann, G. Kullmer, and H. A. Richard, "Stiffness optimization and reliable design of a hip implant by using the potential of additive manufacturing processes," *Biomed Eng Online*, vol. 21, no. 1, Dec. 2022, doi: 10.1186/s12938-022-00990-z.
- [50] R. M. Cowie and L. M. Jennings, "Third body damage and wear in arthroplasty bearing materials: A review of laboratory methods," Dec. 01, 2021, *Elsevier Ltd*. doi: 10.1016/j.bbiosy.2021.100028.
- [51] A. Fouly, M. Taha, T. Albahkali, M. A. Shar, H. S. Abdo, and A. Nabhan, "Developing Artificial Intelligence Models for Predicting the Tribo-Mechanical Properties of HDPE Nanocomposite Used in Artificial Hip Joints," *IEEE Access*, vol. 12, pp. 14787–14799, 2024, doi: 10.1109/ACCESS.2024.3352448.
- [52] D. Milone, G. Risitano, A. Pistone, D. Crisafulli, and F. Alberti, "A New Approach for the Tribological and Mechanical Characterization of a Hip Prosthesis Trough a Numerical Model Based on Artificial Intelligence Algorithms and Humanoid Multibody Model," *Lubricants*, vol. 10, no. 7, Jul. 2022, doi: 10.3390/lubricants10070160.
- [53] Y. Liu *et al.*, "Self-powered artificial joint wear debris sensor based on triboelectric nanogenerator," *Nano Energy*, vol. 85, Jul. 2021, doi: 10.1016/j.nanoen.2021.105967.